

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
 Data File : VX035845.D
 Acq On : 24 May 2023 18:44
 Operator : JC/MD
 Sample : 02911-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01S-20230522

Quant Time: May 25 05:34:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	290094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	123670	42.928	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.860%
35) Dibromofluoromethane	5.385	113	93303	47.692	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.380%
50) Toluene-d8	8.646	98	320875	44.397	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.800%#
62) 4-Bromofluorobenzene	11.079	95	125428	43.102	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.200%
Target Compounds						
					Qvalue	
16) Acetone	2.410	43	1883	1.521	ug/l #	65
20) Methylene Chloride	2.794	84	498	0.222	ug/l #	86
29) Tetrahydrofuran	5.025	42	38663	35.888	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
Data File : VX035845.D
Acq On : 24 May 2023 18:44
Operator : JC/MD
Sample : 02911-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01S-20230522

Quant Time: May 25 05:34:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

