

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052424\
 Data File : VX041643.D
 Acq On : 24 May 2024 10:23
 Operator : JC/MD
 Sample : VX0524WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0524WBL01

Quant Time: May 25 04:19:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144391	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	215423	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95131	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81912	52.110	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.220%	
35) Dibromofluoromethane	5.385	113	67945	46.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.080%	
50) Toluene-d8	8.647	98	247147	53.488	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.980%	
62) 4-Bromofluorobenzene	11.079	95	89665	50.130	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052424\
Data File : VX041643.D
Acq On : 24 May 2024 10:23
Operator : JC/MD
Sample : VX0524WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0524WBL01

Quant Time: May 25 04:19:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

