

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052424\
 Data File : VX041647.D
 Acq On : 24 May 2024 12:11
 Operator : JC/MD
 Sample : P2602-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-052224

Quant Time: May 25 04:22:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143007	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	210975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93100	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80471	51.689	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.380%
35) Dibromofluoromethane	5.385	113	66379	46.428	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	8.647	98	243925	53.904	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.800%
62) 4-Bromofluorobenzene	11.079	95	88760	50.670	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.340%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	3489	1.196	ug/l	# 88
20) Methylene Chloride	2.788	84	814	0.602	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052424\
Data File : VX041647.D
Acq On : 24 May 2024 12:11
Operator : JC/MD
Sample : P2602-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-052224

Quant Time: May 25 04:22:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

