

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
 Data File : VX022166.D
 Acq On : 25 May 2021 10:56
 Operator : JC/MD
 Sample : VX0525MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0525MBL01

Quant Time: May 26 01:53:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	74576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	151371	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47847	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	59124	51.39	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.78%
35) Dibromofluoromethane	5.391	113	46748	48.70	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.40%
50) Toluene-d8	8.653	98	182497	49.35	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.70%
62) 4-Bromofluorobenzene	11.085	95	61158	44.75	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
Data File : VX022166.D
Acq On : 25 May 2021 10:56
Operator : JC/MD
Sample : VX0525MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0525MBL01

Quant Time: May 26 01:53:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

