

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
 Data File : VX022170.D
 Acq On : 25 May 2021 12:38
 Operator : JC/MD
 Sample : M2490-02
 Misc : 4.86g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-11986

Quant Time: May 26 01:55:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	69151	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	141365	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	124057	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	52406	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55259	51.80	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.60%
35) Dibromofluoromethane	5.397	113	42251	47.13	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.26%
50) Toluene-d8	8.653	98	172019	49.81	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.62%
62) 4-Bromofluorobenzene	11.085	95	61275	48.01	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.02%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
Data File : VX022170.D
Acq On : 25 May 2021 12:38
Operator : JC/MD
Sample : M2490-02
Misc : 4.86g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-11986

Quant Time: May 26 01:55:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

