

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
 Data File : VX022173.D
 Acq On : 25 May 2021 13:49
 Operator : JC/MD
 Sample : M2504-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Manual Integrations
 APPROVED

MMDadoda
 5/27/2021 10:30:16 AM

Quant Time: May 26 01:56:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	144655	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	123889	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47233	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	56356	51.84	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.68%
35) Dibromofluoromethane	5.397	113	44926	48.97	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.94%
50) Toluene-d8	8.653	98	175215	49.58	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.16%
62) 4-Bromofluorobenzene	11.085	95	59552	45.59	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.18%
Target Compounds						Qvalue
11) Tert butyl alcohol	2.971	59	2290	9.21	ug/l	98
16) Acetone	2.392	43	2575m	5.44	ug/l	
19) Methyl tert-butyl Ether	3.123	73	1280	0.46	ug/l #	74
83) tert-Butylbenzene	11.719	119	2803	0.88	ug/l	81
84) 1,2,4-Trimethylbenzene	11.756	105	1091	0.33	ug/l	100
85) sec-Butylbenzene	11.890	105	2058	0.51	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
Data File : VX022173.D
Acq On : 25 May 2021 13:49
Operator : JC/MD
Sample : M2504-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Manual Integrations
APPROVED

MMDadoda
5/27/2021 10:30:16 AM

Quant Time: May 26 01:56:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

