

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
 Data File : VX022174.D
 Acq On : 25 May 2021 14:12
 Operator : JC/MD
 Sample : M2504-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-104

Quant Time: May 26 01:56:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	68904	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	144726	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125598	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47905	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55504	52.22	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.44%
35) Dibromofluoromethane	5.397	113	44575	48.57	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.14%
50) Toluene-d8	8.653	98	173872	49.18	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.36%
62) 4-Bromofluorobenzene	11.085	95	59999	45.91	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.82%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
Data File : VX022174.D
Acq On : 25 May 2021 14:12
Operator : JC/MD
Sample : M2504-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-104

Quant Time: May 26 01:56:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

