

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
 Data File : VX022165.D
 Acq On : 25 May 2021 10:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 26 01:53:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.608	0.600	1.3	89	0.00
3 P	Chloromethane	0.656	0.605	7.8	84	0.00
4 C	Vinyl Chloride	0.694	0.692	0.3#	90	0.00
5 T	Bromomethane	0.404	0.364	9.9	89	-0.01
6 T	Chloroethane	0.413	0.421	-1.9	91	0.00
7 T	Trichlorofluoromethane	0.998	1.017	-1.9	93	0.00
8 T	Diethyl Ether	0.429	0.435	-1.4	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.592	0.605	-2.2	95	0.00
10 T	Methyl Iodide	0.637	0.574	9.9	81	0.00
11 T	Tert butyl alcohol	0.176	0.187	-6.3	102	-0.01
12 CM	1,1-Dichloroethene	0.573	0.581	-1.4#	94	0.00
13 T	Acrolein	0.117	0.113	3.4	98	0.00
14 T	Allyl chloride	1.053	1.093	-3.8	96	0.00
15 T	Acrylonitrile	0.347	0.383	-10.4	101	0.00
16 T	Acetone	0.336	0.366	-8.9	106	0.00
17 T	Carbon Disulfide	1.501	1.358	9.5	88	0.00
18 T	Methyl Acetate	0.893	0.967	-8.3	104	0.00
19 T	Methyl tert-butyl Ether	1.963	2.087	-6.3	98	0.00
20 T	Methylene Chloride	0.746	0.722	3.2	97	0.00
21 T	trans-1,2-Dichloroethene	0.623	0.644	-3.4	93	0.00
22 T	Diisopropyl ether	2.027	2.153	-6.2	99	0.00
23 T	Vinyl Acetate	1.952	2.107	-7.9	99	0.00
24 P	1,1-Dichloroethane	1.190	1.236	-3.9	96	0.00
25 T	2-Butanone	0.543	0.603	-11.0	102	0.00
26 T	2,2-Dichloropropane	1.056	1.080	-2.3	95	0.00
27 T	cis-1,2-Dichloroethene	0.758	0.792	-4.5	94	0.00
28 T	Bromochloromethane	0.577	0.583	-1.0	97	0.00
29 T	Tetrahydrofuran	0.323	0.345	-6.8	100	0.00
30 C	Chloroform	1.229	1.274	-3.7#	96	0.00
31 T	Cyclohexane	0.969	0.951	1.9	92	-0.01
32 T	1,1,1-Trichloroethane	1.050	1.073	-2.2	94	0.00
33 S	1,2-Dichloroethane-d4	0.771	0.786	-1.9	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.317	0.315	0.6	102	0.00
36 T	1,1-Dichloropropene	0.462	0.477	-3.2	96	0.00
37 T	Ethyl Acetate	0.567	0.592	-4.4	101	0.00
38 T	Carbon Tetrachloride	0.452	0.467	-3.3	92	-0.01
39 T	Methylcyclohexane	0.524	0.515	1.7	94	0.00
40 TM	Benzene	1.423	1.464	-2.9	96	0.00
41 T	Methacrylonitrile	0.303	0.326	-7.6	102	0.00
42 TM	1,2-Dichloroethane	0.498	0.519	-4.2	97	0.00
43 T	Isopropyl Acetate	0.927	0.974	-5.1	99	0.00
44 TM	Trichloroethene	0.354	0.359	-1.4	95	0.00
45 C	1,2-Dichloropropane	0.363	0.391	-7.7#	96	0.00
46 T	Dibromomethane	0.246	0.263	-6.9	96	0.00
47 T	Bromodichloromethane	0.493	0.525	-6.5	96	0.00
48 T	Methyl methacrylate	0.427	0.456	-6.8	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
 Data File : VX022165.D
 Acq On : 25 May 2021 10:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 26 01:53:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	103	-0.02
50 S	Toluene-d8	1.221	1.200	1.7	101	0.00
51 T	4-Methyl-2-Pentanone	0.555	0.610	-9.9	100	0.00
52 CM	Toluene	0.901	0.927	-2.9#	96	0.00
53 T	t-1,3-Dichloropropene	0.569	0.606	-6.5	98	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.643	-5.2	97	0.00
55 T	1,1,2-Trichloroethane	0.366	0.389	-6.3	98	0.00
56 T	Ethyl methacrylate	0.591	0.648	-9.6	98	0.00
57 T	1,3-Dichloropropane	0.624	0.669	-7.2	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.320	0.335	-4.7	100	0.00
59 T	2-Hexanone	0.418	0.473	-13.2	103	0.00
60 T	Dibromochloromethane	0.356	0.387	-8.7	97	0.00
61 T	1,2-Dibromoethane	0.364	0.398	-9.3	99	0.00
62 S	4-Bromofluorobenzene	0.451	0.463	-2.7	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.328	0.342	-4.3	96	0.00
65 PM	Chlorobenzene	1.041	1.078	-3.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.362	0.382	-5.5	94	0.00
67 C	Ethyl Benzene	1.915	1.972	-3.0#	96	0.00
68 T	m/p-Xylenes	0.724	0.747	-3.2	97	0.00
69 T	o-Xylene	0.704	0.741	-5.3	98	0.00
70 T	Styrene	1.153	1.264	-9.6	99	0.00
71 P	Bromoform	0.250	0.273	-9.2	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	4.232	4.249	-0.4	96	0.00
74 T	N-amyl acetate	2.009	2.198	-9.4	103	0.00
75 P	1,1,2,2-Tetrachloroethane	1.403	1.543	-10.0	105	0.00
76 T	1,2,3-Trichloropropane	1.234	1.395	-13.0	104	0.00
77 T	Bromobenzene	0.897	0.920	-2.6	96	0.00
78 T	n-propylbenzene	4.957	5.216	-5.2	99	0.00
79 T	2-Chlorotoluene	3.006	3.032	-0.9	98	0.00
80 T	1,3,5-Trimethylbenzene	3.507	3.646	-4.0	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.521	0.533	-2.3	94	0.00
82 T	4-Chlorotoluene	3.464	3.529	-1.9	99	0.00
83 T	tert-Butylbenzene	3.364	3.398	-1.0	95	0.00
84 T	1,2,4-Trimethylbenzene	3.530	3.732	-5.7	99	0.00
85 T	sec-Butylbenzene	4.275	4.474	-4.7	97	0.00
86 T	p-Isopropyltoluene	3.610	3.805	-5.4	99	0.00
87 T	1,3-Dichlorobenzene	1.730	1.781	-2.9	98	0.00
88 T	1,4-Dichlorobenzene	1.788	1.817	-1.6	99	0.00
89 T	n-Butylbenzene	3.340	3.606	-8.0	99	0.00
90 T	Hexachloroethane	0.597	0.614	-2.8	93	0.00
91 T	1,2-Dichlorobenzene	1.662	1.751	-5.4	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.300	0.323	-7.7	101	0.00
93 T	1,2,4-Trichlorobenzene	1.003	1.091	-8.8	104	0.00
94 T	Hexachlorobutadiene	0.391	0.411	-5.1	99	0.00
95 T	Naphthalene	3.910	4.346	-11.2	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
Data File : VX022165.D
Acq On : 25 May 2021 10:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 26 01:53:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.011	1.082	-7.0	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6