



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052521\
 Data File : VX022164.D
 Acq On : 25 May 2021 09:19
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Title : SW846 8260
 Last Update : Wed May 19 14:23:10 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	4456	PASS
75	95	30	60	50.3	12332	PASS
95	95	100	100	100.0	24512	PASS
96	95	5	9	7.2	1754	PASS
173	174	0.00	2	0.5	85	PASS
174	95	50	100	64.5	15810	PASS
175	174	5	9	7.1	1129	PASS
176	174	95	101	99.9	15801	PASS
177	176	5	9	6.6	1042	PASS