

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028952.D
 Acq On : 25 May 2022 15:50
 Operator : JC/MD
 Sample : N3030-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 61BR-20220524

Quant Time: May 26 05:45:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	324965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	585216	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	560517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	261730	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	224122	50.313	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.620%	
35) Dibromofluoromethane	5.385	113	196091	51.142	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.280%	
50) Toluene-d8	8.653	98	692447	47.901	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.800%	
62) 4-Bromofluorobenzene	11.079	95	273077	50.576	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.160%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	1775	0.411	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	25889	5.851	ug/l	89
44) Trichloroethene	7.129	130	40600	9.045	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028952.D
Acq On : 25 May 2022 15:50
Operator : JC/MD
Sample : N3030-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
61BR-20220524

Quant Time: May 26 05:45:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

