

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028953.D
 Acq On : 25 May 2022 16:14
 Operator : JC/MD
 Sample : N3030-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB05-20220524

Quant Time: May 26 05:45:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	319567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	572239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	550788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	258941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	223600	51.044	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.080%	
35) Dibromofluoromethane	5.391	113	195660	52.187	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.380%	
50) Toluene-d8	8.653	98	681831	48.236	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.480%	
62) 4-Bromofluorobenzene	11.079	95	270535	51.241	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.480%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	20479	10.893	ug/l	90
20) Methylene Chloride	2.794	84	4688	1.189	ug/l #	92
25) 2-Butanone	4.574	43	3959	1.342	ug/l #	85
27) cis-1,2-Dichloroethene	4.501	96	4253	0.977	ug/l	99
29) Tetrahydrofuran	5.019	42	848	0.442	ug/l #	59
44) Trichloroethene	7.135	130	28269	6.441	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028953.D
Acq On : 25 May 2022 16:14
Operator : JC/MD
Sample : N3030-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB05-20220524

Quant Time: May 26 05:45:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

