

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028955.D
 Acq On : 25 May 2022 17:01
 Operator : JC/MD
 Sample : N3034-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FD-1

Quant Time: May 26 05:45:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	314591	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	565187	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	551197	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	270739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	220838	51.211	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.420%
35) Dibromofluoromethane	5.385	113	192671	52.031	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.060%
50) Toluene-d8	8.653	98	679272	48.655	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.085	95	272065	52.174	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.340%
Target Compounds						
					Qvalue	
3) Chloromethane	1.294	50	929	0.239	ug/l	99
16) Acetone	2.379	43	26198	14.155	ug/l	96
25) 2-Butanone	4.574	43	2390	0.823	ug/l #	86
27) cis-1,2-Dichloroethene	4.501	96	2535	0.592	ug/l	80
44) Trichloroethene	7.129	130	15473	3.569	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028955.D
Acq On : 25 May 2022 17:01
Operator : JC/MD
Sample : N3034-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FD-1

Quant Time: May 26 05:45:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

