

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028958.D
 Acq On : 25 May 2022 18:11
 Operator : JC/MD
 Sample : N2981-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TGMP-3

Quant Time: May 26 05:46:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	312735	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	570458	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	550416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	257407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	221099	51.576	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.160%
35) Dibromofluoromethane	5.391	113	191653	51.278	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.560%
50) Toluene-d8	8.653	98	682789	48.455	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.900%
62) 4-Bromofluorobenzene	11.085	95	268767	51.065	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.140%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.946	59	17080	11.151	ug/l	# 74
16) Acetone	2.379	43	12775	6.944	ug/l	98
20) Methylene Chloride	2.788	84	51560	13.358	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	1230	0.289	ug/l	73
44) Trichloroethene	7.129	130	8861	2.025	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028958.D
Acq On : 25 May 2022 18:11
Operator : JC/MD
Sample : N2981-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TGMP-3

Quant Time: May 26 05:46:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

