

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028959.D
 Acq On : 25 May 2022 18:34
 Operator : JC/MD
 Sample : N2981-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TGMP-4

Quant Time: May 26 05:46:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	322667	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	583127	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	567124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	267242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	227045	51.333	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.660%	
35) Dibromofluoromethane	5.391	113	197081	51.584	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.160%	
50) Toluene-d8	8.653	98	697125	48.397	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.800%	
62) 4-Bromofluorobenzene	11.079	95	275216	51.155	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.300%	
Target Compounds						
						Qvalue
7) Trichlorofluoromethane	1.892	101	2474	0.362	ug/l	84
16) Acetone	2.380	43	16550	8.719	ug/l	99
20) Methylene Chloride	2.788	84	1044	0.262	ug/l	# 88
25) 2-Butanone	4.568	43	2419	0.812	ug/l	99
27) cis-1,2-Dichloroethene	4.507	96	1118	0.254	ug/l	60
44) Trichloroethene	7.129	130	8511	1.903	ug/l	88
95) Naphthalene	13.780	128	9545	0.503	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028959.D
Acq On : 25 May 2022 18:34
Operator : JC/MD
Sample : N2981-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TGMP-4

Quant Time: May 26 05:46:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

