

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028961.D
 Acq On : 25 May 2022 19:21
 Operator : JC/MD
 Sample : N2990-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB-220517

Quant Time: May 26 05:46:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	323556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	578261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	564334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	264570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	225255	50.788	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.580%
35) Dibromofluoromethane	5.385	113	193614	51.103	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.200%
50) Toluene-d8	8.653	98	688848	48.225	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.460%
62) 4-Bromofluorobenzene	11.079	95	274723	51.493	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.980%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	3353	1.762	ug/l	95
29) Tetrahydrofuran	5.019	42	8153	4.198	ug/l	94
44) Trichloroethene	7.129	130	5513	1.243	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028961.D
Acq On : 25 May 2022 19:21
Operator : JC/MD
Sample : N2990-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB-220517

Quant Time: May 26 05:46:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

