

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028967.D
 Acq On : 25 May 2022 22:49
 Operator : JC/MD
 Sample : VX0525WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 VX0525WBL02

Quant Time: May 26 05:49:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.556	168	327239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	587985	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	567742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265902	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	223841	49.901	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.800%
35) Dibromofluoromethane	5.391	113	195729	50.807	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.620%
50) Toluene-d8	8.653	98	700104	48.203	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.400%
62) 4-Bromofluorobenzene	11.079	95	279364	51.497	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028967.D
 Acq On : 25 May 2022 22:49
 Operator : JC/MD
 Sample : VX0525WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0525WBL02

Quant Time: May 26 05:49:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

