

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028972.D
 Acq On : 26 May 2022 00:46
 Operator : JC/MD
 Sample : N3003-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 35S-20220523

Quant Time: May 26 05:50:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	336950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	613136	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	592969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	281741	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	229518	49.692	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.380%
35) Dibromofluoromethane	5.385	113	200735	49.969	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.940%
50) Toluene-d8	8.653	98	728462	48.098	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.200%
62) 4-Bromofluorobenzene	11.079	95	292734	51.748	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.500%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3505	1.768	ug/l	98
20) Methylene Chloride	2.788	84	1156	0.278	ug/l	95
44) Trichloroethene	7.122	130	2011	0.428	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028972.D
Acq On : 26 May 2022 00:46
Operator : JC/MD
Sample : N3003-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
35S-20220523

Quant Time: May 26 05:50:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

