

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052522\  
 Data File : VX028978.D  
 Acq On : 26 May 2022 03:05  
 Operator : JC/MD  
 Sample : N3002-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 45 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-11

Quant Time: May 26 05:53:04 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051222W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 12 14:58:53 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 321913   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 577679   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 549983   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 251957   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 220727   | 50.021   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | = 100.040% |          |
| 35) Dibromofluoromethane    | 5.391  | 113   | 192147   | 50.767   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | = 101.540% |          |
| 50) Toluene-d8              | 8.653  | 98    | 685180   | 48.017   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | = 96.040%  |          |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 266454   | 49.993   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | = 99.980%  |          |
| Target Compounds            |        |       |          |          |            |          |
|                             |        |       |          |          | Qvalue     |          |
| 29) Tetrahydrofuran         | 5.026  | 42    | 1333     | 0.690    | ug/l #     | 47       |
| 44) Trichloroethene         | 7.129  | 130   | 1962     | 0.443    | ug/l       | 98       |
| -----                       |        |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052522\  
Data File : VX028978.D  
Acq On : 26 May 2022 03:05  
Operator : JC/MD  
Sample : N3002-03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 45 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-11

Quant Time: May 26 05:53:04 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X051222W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 12 14:58:53 2022  
Response via : Initial Calibration

