

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028984.D
 Acq On : 26 May 2022 05:25
 Operator : JC/MD
 Sample : N3003-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 35-MW-2-20220523

Quant Time: May 26 08:17:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	337964	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	600599	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	575997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	267949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	223322	48.206	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.420%
35) Dibromofluoromethane	5.391	113	199698	50.749	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.500%
50) Toluene-d8	8.653	98	712442	48.022	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.040%
62) 4-Bromofluorobenzene	11.079	95	278523	50.263	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.520%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	3916	1.970	ug/l	98
44) Trichloroethene	7.129	130	3478	0.755	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028984.D
Acq On : 26 May 2022 05:25
Operator : JC/MD
Sample : N3003-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
35-MW-2-20220523

Quant Time: May 26 08:17:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

