

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
 Data File : VX028985.D
 Acq On : 26 May 2022 05:48
 Operator : JC/MD
 Sample : N3040-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW205S1-20220524

Quant Time: May 26 08:17:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	338216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	606465	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	586135	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	272648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	226999	48.963	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.920%
35) Dibromofluoromethane	5.391	113	200508	50.462	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.920%
50) Toluene-d8	8.653	98	722721	48.243	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.480%
62) 4-Bromofluorobenzene	11.079	95	286342	51.175	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.340%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5221	2.624	ug/l	97
24) 1,1-Dichloroethane	3.611	63	4099	0.588	ug/l	99
44) Trichloroethene	7.129	130	12402	2.666	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052522\
Data File : VX028985.D
Acq On : 26 May 2022 05:48
Operator : JC/MD
Sample : N3040-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW205S1-20220524

Quant Time: May 26 08:17:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

