

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035853.D
 Acq On : 25 May 2023 14:08
 Operator : JC/MD
 Sample : VX0525MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0525MBL01

Quant Time: May 26 07:23:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	209643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	358719	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	324184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	161108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	147800	40.412	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	80.820%
35) Dibromofluoromethane	5.379	113	114694	47.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.820%
50) Toluene-d8	8.647	98	394884	44.185	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.360%#
62) 4-Bromofluorobenzene	11.079	95	161500	44.880	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035853.D
Acq On : 25 May 2023 14:08
Operator : JC/MD
Sample : VX0525MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0525MBL01

Quant Time: May 26 07:23:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

