

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035858.D
 Acq On : 25 May 2023 17:08
 Operator : JC/MD
 Sample : 02908-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 26 07:24:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	207398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	368426	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	347777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167194	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	159691	44.136	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.280%
35) Dibromofluoromethane	5.385	113	128854	51.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.720%
50) Toluene-d8	8.647	98	408176	44.469	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.940%#
62) 4-Bromofluorobenzene	11.079	95	190139	51.447	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035858.D
Acq On : 25 May 2023 17:08
Operator : JC/MD
Sample : 02908-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: May 26 07:24:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

