

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035860.D
 Acq On : 25 May 2023 17:54
 Operator : JC/MD
 Sample : 02881-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: May 26 07:25:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	247209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	477739	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	430093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	209227	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	215060	49.867	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.740%
35) Dibromofluoromethane	5.385	113	159970	49.651	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.300%
50) Toluene-d8	8.647	98	527238	44.297	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.600%#
62) 4-Bromofluorobenzene	11.079	95	223121	46.557	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.120%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035860.D
Acq On : 25 May 2023 17:54
Operator : JC/MD
Sample : 02881-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: May 26 07:25:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

