

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035861.D
 Acq On : 25 May 2023 18:17
 Operator : JC/MD
 Sample : 02881-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: May 26 07:25:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	265830	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	497016	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	419295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	212820	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	203570	43.896	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.800%
35) Dibromofluoromethane	5.385	113	155968	46.532	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	8.647	98	553405	44.692	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.380%#
62) 4-Bromofluorobenzene	11.079	95	222192	44.565	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.140%
Target Compounds						
					Qvalue	
3) Chloromethane	1.301	50	1074	0.255	ug/l #	88
20) Methylene Chloride	2.788	84	4846	1.342	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035861.D
Acq On : 25 May 2023 18:17
Operator : JC/MD
Sample : 02881-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: May 26 07:25:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

