

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035862.D
 Acq On : 25 May 2023 18:41
 Operator : JC/MD
 Sample : 02881-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 26 07:25:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	255616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	474108	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	430551	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	206828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	238876	53.568	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	107.140%	
35) Dibromofluoromethane	5.385	113	148503	46.445	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	92.900%	
50) Toluene-d8	8.647	98	535692	45.352	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	90.700%#	
62) 4-Bromofluorobenzene	11.079	95	233298	49.054	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	98.100%	
Target Compounds						
20) Methylene Chloride	2.788	84	1025	0.295	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035862.D
Acq On : 25 May 2023 18:41
Operator : JC/MD
Sample : 02881-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 26 07:25:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

