

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035866.D
 Acq On : 25 May 2023 20:14
 Operator : JC/MD
 Sample : 02911-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT172S1-20230523RE

Quant Time: May 26 07:25:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	202435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	353074	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	149972	42.466	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.940%
35) Dibromofluoromethane	5.385	113	115893	48.672	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.340%
50) Toluene-d8	8.647	98	386147	43.898	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.800%#
62) 4-Bromofluorobenzene	11.079	95	156374	44.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.300%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	2069	0.872	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	748	0.252	ug/l	55
30) Chloroform	5.105	83	1292	0.249	ug/l	95
44) Trichloroethene	7.123	130	64978	24.372	ug/l	98
64) Tetrachloroethene	9.275	164	2378	1.118	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035866.D
Acq On : 25 May 2023 20:14
Operator : JC/MD
Sample : 02911-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT172S1-20230523RE

Quant Time: May 26 07:25:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

