

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035868.D
 Acq On : 25 May 2023 21:00
 Operator : JC/MD
 Sample : 02911-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT150S1-20230522

Quant Time: May 26 07:26:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	181030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	322121	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	136211	43.130	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.260%
35) Dibromofluoromethane	5.385	113	105743	48.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.360%
50) Toluene-d8	8.647	98	355718	44.325	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.640%#
62) 4-Bromofluorobenzene	11.079	95	139957	43.313	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.620%
Target Compounds						
					Qvalue	
30) Chloroform	5.086	83	1429	0.308	ug/l	94
44) Trichloroethene	7.123	130	23588	9.698	ug/l	98
64) Tetrachloroethene	9.275	164	2055	1.063	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035868.D
Acq On : 25 May 2023 21:00
Operator : JC/MD
Sample : 02911-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT150S1-20230522

Quant Time: May 26 07:26:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

