

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035870.D
 Acq On : 25 May 2023 21:47
 Operator : JC/MD
 Sample : 02911-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20230522

Quant Time: May 26 07:26:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	184924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	330263	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	293539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141299	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	140651	43.598	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.200%
35) Dibromofluoromethane	5.385	113	109481	49.154	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	8.647	98	365889	44.468	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.940%#
62) 4-Bromofluorobenzene	11.079	95	148009	44.675	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.360%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	2072	0.860	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.331	101	40598	18.736	ug/l	98
12) 1,1-Dichloroethene	2.319	96	6081	2.953	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	8476	3.125	ug/l	86
30) Chloroform	5.093	83	6695	1.414	ug/l	91
38) Carbon Tetrachloride	5.678	117	4391	1.318	ug/l	90
44) Trichloroethene	7.123	130	748273	300.051	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	1985	0.814	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035870.D
Acq On : 25 May 2023 21:47
Operator : JC/MD
Sample : 02911-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20230522

Quant Time: May 26 07:26:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

