

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
 Data File : VX035873.D
 Acq On : 25 May 2023 22:57
 Operator : JC/MD
 Sample : 02911-19 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D2-20230522

Quant Time: May 26 07:26:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	203859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	364605	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	327634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	153642	43.202	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.400%
35) Dibromofluoromethane	5.385	113	118896	48.354	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.700%
50) Toluene-d8	8.647	98	402067	44.262	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.520%#
62) 4-Bromofluorobenzene	11.079	95	159597	43.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.280%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	9857	4.127	ug/l	97
12) 1,1-Dichloroethene	2.319	96	5439	2.396	ug/l	82
20) Methylene Chloride	2.788	84	1232	0.445	ug/l	92
27) cis-1,2-Dichloroethene	4.501	96	876	0.293	ug/l	58
44) Trichloroethene	7.123	130	224741	81.631	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052523\
Data File : VX035873.D
Acq On : 25 May 2023 22:57
Operator : JC/MD
Sample : 02911-19 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D2-20230522

Quant Time: May 26 07:26:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

