

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052620\
 Data File : VX016433.D
 Acq On : 26 May 2020 18:27
 Operator : JC/SP
 Sample : PB129074TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129074TB

Quant Time: May 27 05:53:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 26 16:32:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	109068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	196815	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	203545	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	94833	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	75511	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
35) Dibromofluoromethane	5.47	113	58004	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
50) Toluene-d8	8.70	98	257902	55.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.24%	
62) 4-Bromofluorobenzene	11.12	95	98012	56.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.92%	
Target Compounds						
16) Acetone	2.42	43	8974	15.948	ug/l	91
18) Methyl Acetate	2.75	43	1495	1.089	ug/l #	82
20) Methylene Chloride	2.84	84	2416	1.990	ug/l #	85
43) Isopropyl Acetate	6.42	43	37152	12.011	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052620\
Data File : VX016433.D
Acq On : 26 May 2020 18:27
Operator : JC/SP
Sample : PB129074TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129074TB

Quant Time: May 27 05:53:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
Quant Title : SW846 8260
QLast Update : Tue May 26 16:32:50 2020
Response via : Initial Calibration

