

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052620\
 Data File : VX016437.D
 Acq On : 26 May 2020 19:58
 Operator : JC/SP
 Sample : PB129074ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129074ZHE#04

Quant Time: May 27 06:16:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 26 16:32:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	108234	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	201153	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	222003	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	104400	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	79155	56.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.92%	
35) Dibromofluoromethane	5.47	113	61438	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
50) Toluene-d8	8.70	98	272405	57.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.96%	
62) 4-Bromofluorobenzene	11.12	95	109017	61.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.96%	
Target Compounds						
16) Acetone	2.42	43	8914	15.963	ug/l	98
18) Methyl Acetate	2.75	43	1848	1.357	ug/l	98
20) Methylene Chloride	2.84	84	2013	1.671	ug/l	95
43) Isopropyl Acetate	6.42	43	38948	12.320	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052620\
Data File : VX016437.D
Acq On : 26 May 2020 19:58
Operator : JC/SP
Sample : PB129074ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129074ZHE#04

Quant Time: May 27 06:16:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
Quant Title : SW846 8260
QLast Update : Tue May 26 16:32:50 2020
Response via : Initial Calibration

