

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052620\
 Data File : VX016440.D
 Acq On : 26 May 2020 21:06
 Operator : JC/SP
 Sample : L2500-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-03-052220

Quant Time: May 27 06:26:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 26 16:32:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	105978	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	196681	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	204764	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	96258	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	76017	55.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.76%	
35) Dibromofluoromethane	5.47	113	59606	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	8.70	98	254818	54.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.98%	
62) 4-Bromofluorobenzene	11.12	95	99472	57.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.68%	
Target Compounds						
16) Acetone	2.42	43	16138	29.515	ug/l	99
20) Methylene Chloride	2.83	84	3261	2.765	ug/l #	88
43) Isopropyl Acetate	6.42	43	22712	7.348	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052620\
Data File : VX016440.D
Acq On : 26 May 2020 21:06
Operator : JC/SP
Sample : L2500-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
OK-03-052220

Quant Time: May 27 06:26:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
Quant Title : SW846 8260
QLast Update : Tue May 26 16:32:50 2020
Response via : Initial Calibration

