

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052620\
 Data File : VX016442.D
 Acq On : 26 May 2020 21:51
 Operator : JC/SP
 Sample : L2757-02 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0AP4

Quant Time: May 27 06:32:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 26 16:32:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	110882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	201350	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	207997	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	104172	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	81802	56.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.90%	
35) Dibromofluoromethane	5.47	113	63002	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	8.70	98	263929	55.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.28%	
62) 4-Bromofluorobenzene	11.12	95	103747	58.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.86%	
Target Compounds						
18) Methyl Acetate	2.75	43	7582	5.434	ug/l	98
52) Toluene	8.77	92	19236	5.627	ug/l	96
67) Ethyl Benzene	10.24	91	9769	1.367	ug/l	97
68) m/p-Xylenes	10.34	106	8854	3.169	ug/l	100
69) o-Xylene	10.68	106	5289	1.992	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052620\
Data File : VX016442.D
Acq On : 26 May 2020 21:51
Operator : JC/SP
Sample : L2757-02 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AP4

Quant Time: May 27 06:32:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
Quant Title : SW846 8260
QLast Update : Tue May 26 16:32:50 2020
Response via : Initial Calibration

