


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX052620\
 Data File : VX016419.D
 Acq On : 26 May 2020 11:53
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052620W.M
 Title : SW846 8260
 Last Update : Tue May 26 16:32:50 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	5761	PASS
75	95	30	60	54.5	15443	PASS
95	95	100	100	100.0	28354	PASS
96	95	5	9	7.1	2005	PASS
173	174	0.00	2	0.5	113	PASS
174	95	50	100	77.8	22069	PASS
175	174	5	9	8.5	1869	PASS
176	174	95	101	97.3	21464	PASS
177	176	5	9	6.5	1404	PASS