

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052621\
 Data File : VX022184.D
 Acq On : 25 May 2021 18:37
 Operator : JC/MD
 Sample : VX0526WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0526WBL01

Quant Time: May 26 02:36:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML052621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 25 17:28:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	157304	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	140061	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	58441	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	60855	50.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.700%
7) Chloroethane-d5	1.666	69	47982	53.090	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.180%
11) 1,1-Dichloroethene-d2	2.306	63	87049	39.694	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.380%
21) 2-Butanone-d5	4.471	46	103504	109.327	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.330%
24) Chloroform-d	5.068	84	111839	50.696	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.400%
26) 1,2-Dichloroethane-d4	5.964	65	75663	52.237	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.480%
32) Benzene-d6	5.982	84	243262	54.102	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.200%
36) 1,2-Dichloropropane-d6	7.312	67	75673	53.702	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.400%
41) Toluene-d8	8.653	98	215567	52.729	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.460%
43) trans-1,3-Dichloroprop...	8.951	79	35247	50.820	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.640%
47) 2-Hexanone-d5	9.390	63	76483	112.250	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.250%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	100264	52.017	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.040%
66) 1,2-Dichlorobenzene-d4	12.323	152	67981	55.719	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052621\
Data File : VX022184.D
Acq On : 25 May 2021 18:37
Operator : JC/MD
Sample : VX0526WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0526WBL01

Quant Time: May 26 02:36:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML052621WMA.M
Quant Title : VOC Analysis
QLast Update : Tue May 25 17:28:06 2021
Response via : Initial Calibration

