

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052621\
 Data File : VX022182.D
 Acq On : 25 May 2021 17:26
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :

Manual Integrations
 APPROVED

Quant Time: May 26 02:35:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML052621WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 25 17:28:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	161676	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	142743	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	61535	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	65156	52.974	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.940%
7) Chloroethane-d5	1.666	69	52134	56.124	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	112.240%
11) 1,1-Dichloroethene-d2	2.306	63	95399	42.325	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.640%
21) 2-Butanone-d5	4.471	46	107837	110.823	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.820%
24) Chloroform-d	5.068	84	118209	52.135	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.260%
26) 1,2-Dichloroethane-d4	5.964	65	77798	52.259	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.520%
32) Benzene-d6	5.983	84	253957	55.419	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.840%
36) 1,2-Dichloropropane-d6	7.318	67	79165	55.124	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.240%
41) Toluene-d8	8.653	98	225447	54.110	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.220%
43) trans-1,3-Dichloroprop...	8.952	79	38232	54.088	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.180%
47) 2-Hexanone-d5	9.391	63	78729	113.375	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.380%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	103677	52.777	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.560%
66) 1,2-Dichlorobenzene-d4	12.323	152	70360	54.770	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052621\
Data File : VX022182.D
Acq On : 25 May 2021 17:26
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Manual Integrations
APPROVED

Quant Time: May 26 02:35:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML052621WMA.M
Quant Title : VOC Analysis
QLast Update : Tue May 25 17:28:06 2021
Response via : Initial Calibration

