

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
 Data File : VX028998.D
 Acq On : 26 May 2022 14:52
 Operator : JC/MD
 Sample : N3030-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 61BR-20220524

Quant Time: May 27 07:33:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	326622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	582420	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	562538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	266031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	217584	48.598	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.200%		
35) Dibromofluoromethane	5.391	113	191393	50.156	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.320%		
50) Toluene-d8	8.653	98	693607	48.211	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.420%		
62) 4-Bromofluorobenzene	11.079	95	276431	51.443	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	102.880%		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	20254	4.554	ug/l	90
44) Trichloroethene	7.135	130	4124	0.923	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
Data File : VX028998.D
Acq On : 26 May 2022 14:52
Operator : JC/MD
Sample : N3030-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
61BR-20220524

Quant Time: May 27 07:33:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

