

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
 Data File : VX029001.D
 Acq On : 26 May 2022 16:02
 Operator : JC/MD
 Sample : N3066-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-MW163S1-TB-20220524

Quant Time: May 27 07:34:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	323416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	575856	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	550939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	257289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	217552	49.072	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.140%
35) Dibromofluoromethane	5.385	113	190948	50.610	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.220%
50) Toluene-d8	8.653	98	687698	48.346	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.700%
62) 4-Bromofluorobenzene	11.079	95	267606	50.368	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.740%
Target Compounds						
16) Acetone	2.374	43	3083	1.620	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
Data File : VX029001.D
Acq On : 26 May 2022 16:02
Operator : JC/MD
Sample : N3066-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-MW163S1-TB-20220524

Quant Time: May 27 07:34:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

