

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
 Data File : VX029003.D
 Acq On : 26 May 2022 16:49
 Operator : JC/MD
 Sample : N3066-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-MW163S1-GW-285-287

Quant Time: May 27 07:34:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	324885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	581339	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	563017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	263558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	219842	49.365	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.720%
35) Dibromofluoromethane	5.391	113	191913	50.386	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.780%
50) Toluene-d8	8.653	98	691798	48.175	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.360%
62) 4-Bromofluorobenzene	11.079	95	276234	51.502	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.000%
Target Compounds						
16) Acetone	2.380	43	10485	5.486	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
Data File : VX029003.D
Acq On : 26 May 2022 16:49
Operator : JC/MD
Sample : N3066-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-MW163S1-GW-285-287

Quant Time: May 27 07:34:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

