

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
 Data File : VX029006.D
 Acq On : 26 May 2022 17:59
 Operator : JC/MD
 Sample : N3073-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 220525029-01-TRIP-BLANK

Quant Time: May 27 07:35:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	321585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	571772	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	546079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	256713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	216889	49.201	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.400%
35) Dibromofluoromethane	5.391	113	191225	51.046	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.100%
50) Toluene-d8	8.653	98	685287	48.520	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	265498	50.328	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.660%
Target Compounds						
16) Acetone	2.380	43	10574	5.589	ug/l	91
18) Methyl Acetate	2.709	43	1308	0.276	ug/l #	79
30) Chloroform	5.105	83	3750	0.522	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
Data File : VX029006.D
Acq On : 26 May 2022 17:59
Operator : JC/MD
Sample : N3073-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
220525029-01-TRIP-BLANK

Quant Time: May 27 07:35:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

