

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
 Data File : VX028988.D
 Acq On : 26 May 2022 10:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 27 07:30:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	126	0.00
2 T	Dichlorodifluoromethane	0.655	0.601	8.2	115	0.00
3 P	Chloromethane	0.617	0.519	15.9	112	0.00
4 C	Vinyl Chloride	0.664	0.585	11.9#	112	0.00
5 T	Bromomethane	0.523	0.542	-3.6	143	0.00
6 T	Chloroethane	0.467	0.388	16.9	114	0.00
7 T	Trichlorofluoromethane	1.059	0.931	12.1	111	0.00
8 T	Diethyl Ether	0.386	0.351	9.1	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.558	-0.2	129	0.00
10 T	Methyl Iodide	0.667	0.644	3.4	118	0.00
11 T	Tert butyl alcohol	0.156	0.144	7.7	130	0.00
12 CM	1,1-Dichloroethene	0.517	0.495	4.3#	122	0.00
13 T	Acrolein	0.112	0.021	81.3#	27#	0.00
14 T	Allyl chloride	0.830	0.805	3.0	123	0.00
15 T	Acrylonitrile	0.330	0.323	2.1	125	0.00
16 T	Acetone	0.294	0.317	-7.8	144	0.00
17 T	Carbon Disulfide	1.341	1.083	19.2	101	0.00
18 T	Methyl Acetate	0.737	0.725	1.6	129	0.00
19 T	Methyl tert-butyl Ether	1.836	1.944	-5.9	133	0.00
20 T	Methylene Chloride	0.617	0.594	3.7	127	0.00
21 T	trans-1,2-Dichloroethene	0.583	0.551	5.5	118	0.00
22 T	Diisopropyl ether	1.774	1.764	0.6	125	0.00
23 T	Vinyl Acetate	1.529	1.497	2.1	120	0.00
24 P	1,1-Dichloroethane	1.031	1.031	0.0	126	0.00
25 T	2-Butanone	0.462	0.447	3.2	124	0.00
26 T	2,2-Dichloropropane	0.808	0.888	-9.9	135	0.00
27 T	cis-1,2-Dichloroethene	0.681	0.689	-1.2	129	0.00
28 T	Bromochloromethane	0.434	0.377	13.1	127	0.00
29 T	Tetrahydrofuran	0.300	0.274	8.7	117	0.00
30 C	Chloroform	1.116	1.136	-1.8#	128	0.00
31 T	Cyclohexane	0.951	0.860	9.6	113	0.00
32 T	1,1,1-Trichloroethane	0.974	1.004	-3.1	127	0.00
33 S	1,2-Dichloroethane-d4	0.685	0.665	2.9	124	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	120	0.00
35 S	Dibromofluoromethane	0.328	0.360	-9.8	131	0.00
36 T	1,1-Dichloropropene	0.492	0.502	-2.0	120	0.00
37 T	Ethyl Acetate	0.514	0.516	-0.4	115	0.00
38 T	Carbon Tetrachloride	0.495	0.558	-12.7	128	0.00
39 T	Methylcyclohexane	0.564	0.606	-7.4	123	0.00
40 TM	Benzene	1.414	1.480	-4.7	124	0.00
41 T	Methacrylonitrile	0.276	0.294	-6.5	128	0.00
42 TM	1,2-Dichloroethane	0.522	0.546	-4.6	124	0.00
43 T	Isopropyl Acetate	0.804	0.840	-4.5	122	0.00
44 TM	Trichloroethene	0.384	0.423	-10.2	128	0.00
45 C	1,2-Dichloropropane	0.353	0.383	-8.5#	129	0.00
46 T	Dibromomethane	0.263	0.277	-5.3	124	0.00
47 T	Bromodichloromethane	0.489	0.558	-14.1	129	0.00
48 T	Methyl methacrylate	0.388	0.414	-6.7	123	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
 Data File : VX028988.D
 Acq On : 26 May 2022 10:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 27 07:30:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	112	0.00
50 S	Toluene-d8	1.235	1.311	-6.2	124	0.00
51 T	4-Methyl-2-Pentanone	0.530	0.544	-2.6	119	0.00
52 CM	Toluene	0.910	0.993	-9.1#	126	0.00
53 T	t-1,3-Dichloropropene	0.496	0.589	-18.7	128	0.00
54 T	cis-1,3-Dichloropropene	0.550	0.641	-16.5	129	0.00
55 T	1,1,2-Trichloroethane	0.371	0.416	-12.1	131	0.00
56 T	Ethyl methacrylate	0.562	0.648	-15.3	130	0.00
57 T	1,3-Dichloropropane	0.626	0.693	-10.7	131	0.00
58 T	2-Chloroethyl Vinyl ether	0.275	0.276	-0.4	121	0.00
59 T	2-Hexanone	0.411	0.423	-2.9	117	0.00
60 T	Dibromochloromethane	0.362	0.446	-23.2	135	0.00
61 T	1,2-Dibromoethane	0.390	0.440	-12.8	130	0.00
62 S	4-Bromofluorobenzene	0.461	0.538	-16.7	134	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	119	0.00
64 T	Tetrachloroethene	0.378	0.417	-10.3	128	0.00
65 PM	Chlorobenzene	1.061	1.178	-11.0	130	0.00
66 T	1,1,1,2-Tetrachloroethane	0.376	0.448	-19.1	137	0.00
67 C	Ethyl Benzene	1.901	2.076	-9.2#	125	0.00
68 T	m/p-Xylenes	0.740	0.815	-10.1	125	0.00
69 T	o-Xylene	0.729	0.813	-11.5	129	0.00
70 T	Styrene	1.183	1.389	-17.4	129	0.00
71 P	Bromoform	0.274	0.348	-27.0#	139	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
73 T	Isopropylbenzene	3.792	4.089	-7.8	129	0.00
74 T	N-amyl acetate	1.491	1.507	-1.1	116	0.00
75 P	1,1,2,2-Tetrachloroethane	1.301	1.354	-4.1	130	0.00
76 T	1,2,3-Trichloropropane	1.171	1.258	-7.4	131	0.00
77 T	Bromobenzene	0.897	0.992	-10.6	132	0.00
78 T	n-propylbenzene	4.386	4.714	-7.5	127	0.00
79 T	2-Chlorotoluene	2.688	2.895	-7.7	131	0.00
80 T	1,3,5-Trimethylbenzene	3.159	3.524	-11.6	131	0.00
81 T	trans-1,4-Dichloro-2-butene	0.357	0.397	-11.2	129	0.00
82 T	4-Chlorotoluene	3.108	3.353	-7.9	129	0.00
83 T	tert-Butylbenzene	3.059	3.565	-16.5	138	0.00
84 T	1,2,4-Trimethylbenzene	3.168	3.518	-11.0	130	0.00
85 T	sec-Butylbenzene	3.773	4.392	-16.4	137	0.00
86 T	p-Isopropyltoluene	3.166	3.782	-19.5	137	0.00
87 T	1,3-Dichlorobenzene	1.760	1.959	-11.3	133	0.00
88 T	1,4-Dichlorobenzene	1.825	1.956	-7.2	131	0.00
89 T	n-Butylbenzene	2.743	3.209	-17.0	132	0.00
90 T	Hexachloroethane	0.512	0.606	-18.4	135	0.00
91 T	1,2-Dichlorobenzene	1.729	1.930	-11.6	133	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.278	0.299	-7.6	128	0.00
93 T	1,2,4-Trichlorobenzene	0.998	1.237	-23.9	145	0.00
94 T	Hexachlorobutadiene	0.406	0.528	-30.0#	153#	0.00
95 T	Naphthalene	3.553	4.116	-15.8	131	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052622\
Data File : VX028988.D
Acq On : 26 May 2022 10:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 27 07:30:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.003	1.210	-20.6	144	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6