

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052720\
 Data File : VX016470.D
 Acq On : 28 May 2020 00:45
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: May 28 03:05:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	394213	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	644834	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	571706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	286924	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	239978	49.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.04%	
35) Dibromofluoromethane	5.47	113	190553	47.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.98%	
50) Toluene-d8	8.70	98	738792	47.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.62%	
62) 4-Bromofluorobenzene	11.12	95	276268	46.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.52%	

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.18	85	177845	45.842	ug/l	98
3) Chloromethane	1.31	50	206854	44.122	ug/l	100
4) Vinyl Chloride	1.39	62	233642	47.125	ug/l	97
5) Bromomethane	1.62	94	92182	38.240	ug/l	98
6) Chloroethane	1.70	64	138955	48.046	ug/l	96
7) Trichlorofluoromethane	1.91	101	287918	44.219	ug/l	97
8) Diethyl Ether	2.17	74	126146	46.462	ug/l	97
9) 1,1,2-Trichlorotrifluoroet	2.36	101	167356	46.701	ug/l	100
10) Methyl Iodide	2.49	142	170061	40.349	ug/l	98
11) Tert butyl alcohol	3.02	59	379423	302.236	ug/l	100
12) 1,1-Dichloroethene	2.36	96	177988	46.923	ug/l	99
13) Acrolein	2.27	56	135005	252.806	ug/l	96
14) Allyl chloride	2.70	41	346821	49.717	ug/l	97
15) Acrylonitrile	3.12	53	683430	265.762	ug/l	99
16) Acetone	2.42	43	606175	282.615	ug/l	100
17) Carbon Disulfide	2.55	76	536152	47.597	ug/l	100
18) Methyl Acetate	2.75	43	292945	53.251	ug/l	99
19) Methyl tert-butyl Ether	3.17	73	658304	48.331	ug/l	99
20) Methylene Chloride	2.83	84	206117	46.447	ug/l	98
21) trans-1,2-Dichloroethene	3.14	96	195120	46.505	ug/l	98
22) Diisopropyl ether	3.82	45	646049	48.209	ug/l	99
23) Vinyl Acetate	3.78	43	2944299	251.009	ug/l	99
24) 1,1-Dichloroethane	3.67	63	362726	47.956	ug/l	99
25) 2-Butanone	4.64	43	997417	274.437	ug/l	100
26) 2,2-Dichloropropane	4.55	77	245032	39.590	ug/l	98
27) cis-1,2-Dichloroethene	4.56	96	223346	46.014	ug/l	100
28) Bromochloromethane	4.98	49	167294	47.596	ug/l	100
29) Tetrahydrofuran	5.09	42	641442	273.186	ug/l	100
30) Chloroform	5.18	83	349639	45.877	ug/l	97
31) Cyclohexane	5.55	56	318147	46.755	ug/l	98
32) 1,1,1-Trichloroethane	5.46	97	313281	45.405	ug/l	100
36) 1,1-Dichloropropene	5.77	75	267746	43.854	ug/l	99
37) Ethyl Acetate	4.79	43	353634	48.245	ug/l	100
38) Carbon Tetrachloride	5.76	117	272569	42.482	ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX052720\
 Data File : VX016470.D
 Acq On : 28 May 2020 00:45
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 VSTDCCC050EC

Quant Time: May 28 03:05:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.44	83	300929	45.379	ug/l	99
40) Benzene	6.11	78	811771	44.648	ug/l	100
41) Methacrylonitrile	5.00	41	188334	46.776	ug/l	99
42) 1,2-Dichloroethane	6.17	62	280726	43.417	ug/l	99
43) Isopropyl Acetate	6.42	43	544001	47.580	ug/l	99
44) Trichloroethene	7.19	130	237385	41.151	ug/l	95
45) 1,2-Dichloropropane	7.49	63	210275	45.430	ug/l	96
46) Dibromomethane	7.64	93	138628	44.591	ug/l	99
47) Bromodichloromethane	7.88	83	284558	44.205	ug/l	99
48) Methyl methacrylate	7.75	41	262840	48.013	ug/l	99
49) 1,4-Dioxane	7.71	88	142218	1110.092	ug/l	99
51) 4-Methyl-2-Pentanone	8.62	43	1773369	249.097	ug/l	100
52) Toluene	8.77	92	505672	44.105	ug/l	100
53) t-1,3-Dichloropropene	9.02	75	327449	44.735	ug/l	99
54) cis-1,3-Dichloropropene	8.42	75	346350	44.703	ug/l	98
55) 1,1,2-Trichloroethane	9.20	97	204887	44.946	ug/l	97
56) Ethyl methacrylate	9.16	69	353747	47.648	ug/l	99
57) 1,3-Dichloropropane	9.35	76	354295	45.280	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.29	63	912901	235.581	ug/l	99
59) 2-Hexanone	9.48	43	1412748	255.180	ug/l	100
60) Dibromochloromethane	9.57	129	227129	44.275	ug/l	100
61) 1,2-Dibromoethane	9.65	107	219539	44.973	ug/l	99
64) Tetrachloroethene	9.32	164	255575	39.957	ug/l	97
65) Chlorobenzene	10.12	112	536155	46.000	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.21	131	203972	45.845	ug/l	98
67) Ethyl Benzene	10.24	91	954405	45.991	ug/l	98
68) m/p-Xylenes	10.34	106	713058	92.675	ug/l	100
69) o-Xylene	10.68	106	346881	46.552	ug/l	97
70) Styrene	10.70	104	609625	47.597	ug/l	99
71) Bromoform	10.84	173	177323	47.158	ug/l #	98
73) Isopropylbenzene	11.00	105	911706	45.860	ug/l	100
74) N-amyl acetate	10.88	43	456418	49.912	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.26	83	269836	54.157	ug/l	100
76) 1,2,3-Trichloropropane	11.28	75	393633	62.415	ug/l	84
77) Bromobenzene	11.24	156	237266	45.503	ug/l	99
78) n-propylbenzene	11.34	91	1035529	46.623	ug/l	100
79) 2-Chlorotoluene	11.40	91	617190	45.524	ug/l	100
80) 1,3,5-Trimethylbenzene	11.49	105	762968	45.962	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.06	75	117933	45.998	ug/l	98
82) 4-Chlorotoluene	11.49	91	732958	45.782	ug/l	99
83) tert-Butylbenzene	11.76	119	682565	48.123	ug/l	98
84) 1,2,4-Trimethylbenzene	11.79	105	774112	46.112	ug/l	100
85) sec-Butylbenzene	11.93	105	840670	47.247	ug/l	99
86) p-Isopropyltoluene	12.05	119	775949	46.555	ug/l	99
87) 1,3-Dichlorobenzene	12.01	146	409121	45.289	ug/l	99
88) 1,4-Dichlorobenzene	12.08	146	407533	44.110	ug/l	98
89) n-Butylbenzene	12.37	91	660710	47.592	ug/l	99
90) Hexachloroethane	12.58	117	135920	47.212	ug/l	97
91) 1,2-Dichlorobenzene	12.38	146	394402	45.063	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	12.98	75	82294	50.992	ug/l	99


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052720\
Data File : VX016470.D
Acq On : 28 May 2020 00:45
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: May 28 03:05:43 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.63	180	274563	47.236	ug/l	99
94) Hexachlorobutadiene	13.77	225	106027	46.459	ug/l	98
95) Naphthalene	13.82	128	984027	49.383	ug/l	99
96) 1,2,3-Trichlorobenzene	14.01	180	273242	47.983	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:168 Resp: 394213

Ion Ratio Lower Upper

168 100

99 56.4 45.2 67.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

150000



100000

50000

0

5.50 5.60 5.70 5.80

Time-->



#2

Dichlorodifluoromethane

Concen: 45.842 ug/l

RT: 1.18 min Scan# 15

Delta R.T. -0.00 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 85 Resp: 177845

Ion Ratio Lower Upper

85 100

87 32.1 15.4 46.2



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

200000

150000


100000


50000

0

1.10 1.15 1.20 1.25 1.30

Time-->





#3

Chloromethane

Concen: 44.122 ug/l

RT: 1.31 min Scan# 37

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 50 Resp: 206854

Ion Ratio Lower Upper

50 100

52 32.6 26.2 39.2



#4

Vinyl Chloride

Concen: 47.125 ug/l

RT: 1.39 min Scan# 50

Delta R.T. -0.00 min

Lab File: VX016470.D




Acq: 28 May 2020 00:45


Tgt Ion: 62 Resp: 233642

Ion Ratio Lower Upper

62 100

64 30.6 25.9 38.9





#5

Bromomethane

Concen: 38.240 ug/l

RT: 1.62 min Scan# 87

Delta R.T. -0.01 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 94 Resp: 92182

Ion Ratio Lower Upper

94 100

96 91.9 75.1 112.7



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0

1.62

80000



60000

40000

20000

0

Time--> 1.55 1.60 1.65



#6

Chloroethane

Concen: 48.046 ug/l

RT: 1.70 min Scan# 101

Delta R.T. -0.00 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 64 Resp: 138955

Ion Ratio Lower Upper

64 100

66 31.0 26.7 40.1



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0


1.70


100000

50000

0

Time--> 1.60 1.70 1.80






#7




Trichlorofluoromethane
Concen: 44.219 ug/l
RT: 1.91 min Scan# 135
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion:101 Resp: 287918
Ion Ratio Lower Upper
101 100
103 65.7 50.9 76.3




Abundance Ion 100.90 (100.60 to 101.60): V
Ion 102.80 (102.50 to 103.50): V





#8


Diethyl Ether
Concen: 46.462 ug/l
RT: 2.17 min Scan# 177
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion: 74 Resp: 126146
Ion Ratio Lower Upper
74 100
45 98.4 47.5 142.5



Abundance Ion 74.00 (73.70 to 74.70): VX0
Ion 45.00 (44.70 to 45.70): VX0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 46.701 ug/l

RT: 2.36 min Scan# 209

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:101 Resp: 167356

Ion Ratio Lower Upper

101 100

85 42.9 34.4 51.6





151 79.3 63.8 95.6



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 40.349 ug/l

RT: 2.49 min Scan# 230

Delta R.T. 0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion:142 Resp: 170061

Ion Ratio Lower Upper

142 100

127 42.3 35.3 52.9




141 14.5 11.4 17.2




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 302.236 ug/l
 RT: 3.02 min Scan# 317
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tot Ion: 59 Resp: 379423
 Ion Ratio Lower Upper
 59 100
 57 10.1 8.2 12.4




#12

1,1-Dichloroethene
 Concen: 46.923 ug/l
 RT: 2.36 min Scan# 208
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion: 96 Resp: 177988
 Ion Ratio Lower Upper
 96 100
 61 159.9 126.6 189.8
 98 65.4 51.3 76.9





#13

Acrolein

Concen: 252.806 ug/l

RT: 2.27 min Scan# 194

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 56 Resp: 135005

Ion Ratio Lower Upper

56 100

55 69.8 58.3 87.5



#14

Allyl chloride

Concen: 49.717 ug/l

RT: 2.70 min Scan# 265

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45




Tgt Ion: 41 Resp: 346821


Ion Ratio Lower Upper

41 100

39 56.6 47.7 71.5

76 31.9 26.6 40.0





#15

Acrylonitrile

Concen: 265.762 ug/l

RT: 3.12 min Scan# 333

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tot Ion: 53 Resp: 683430

Ion Ratio Lower Upper

53 100

52 82.5 66.7 100.1

51 35.1 28.6 43.0



#16

Acetone

Concen: 282.615 ug/l

RT: 2.42 min Scan# 218

Delta R.T. -0.00 min

Lab File: VX016470.D




Acq: 28 May 2020 00:45


Tgt Ion: 43 Resp: 606175

Ion Ratio Lower Upper

43 100

58 32.8 26.1 39.1





#17

Carbon Disulfide

Concen: 47.597 ug/l

RT: 2.55 min Scan# 240

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 76 Resp: 536152

Ion Ratio Lower Upper

76 100

78 9.1 7.4 11.0



#18

Methyl Acetate

Concen: 53.251 ug/l

RT: 2.75 min Scan# 272

Delta R.T. -0.00 min

Lab File: VX016470.D




Acq: 28 May 2020 00:45


Tgt Ion: 43 Resp: 292945

Ion Ratio Lower Upper

43 100

74 24.1 19.7 29.5









#19

Methyl tert-butyl Ether
 Concen: 48.331 ug/l
 RT: 3.17 min Scan# 341
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tot Ion: 73 Resp: 658304
 Ion Ratio Lower Upper
 73 100
 57 21.8 17.8 26.6




#20

Methylene Chloride
 Concen: 46.447 ug/l
 RT: 2.83 min Scan# 286
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion: 84 Resp: 206117
 Ion Ratio Lower Upper
 84 100
 49 122.0 94.5 141.7
 51 38.2 30.2 45.4
 86 62.8 51.1 76.7





#21

trans-1,2-Dichloroethene

Concen: 46.505 ug/l

RT: 3.14 min Scan# 337

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 96 Resp: 195120

Ion Ratio Lower Upper

96 100

61 140.7 115.3 172.9




98 65.2 51.9 77.9



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 48.209 ug/l

RT: 3.82 min Scan# 449

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Tgt Ion: 45 Resp: 646049


Ion Ratio Lower Upper

45 100

43 70.7 57.0 85.6

87 27.9 22.5 33.7

59 11.4 9.1 13.7






Abundance Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 251.009 ug/l

RT: 3.78 min Scan# 442

Delta R.T. -0.01 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 43 Resp: 2944299

Ion Ratio Lower Upper

43 100

86 11.0 8.6 13.0



#24

1,1-Dichloroethane

Concen: 47.956 ug/l

RT: 3.67 min Scan# 424

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45




Tgt Ion: 63 Resp: 362726


Ion Ratio Lower Upper

63 100

98 7.2 3.6 10.9

100 4.0 2.4 7.1





#25

2-Butanone

Concen: 274.437 ug/l

RT: 4.64 min Scan# 582

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 43 Resp: 997417

Ion Ratio Lower Upper

43 100

72 26.4 21.0 31.6



#26

2,2-Dichloropropane

Concen: 39.590 ug/l

RT: 4.55 min Scan# 568

Delta R.T. -0.00 min

Lab File: VX016470.D




Acq: 28 May 2020 00:45


Tgt Ion: 77 Resp: 245032

Ion Ratio Lower Upper

77 100

97 23.7 11.5 34.4









#27

cis-1,2-Dichloroethene
 Concen: 46.014 ug/l
 RT: 4.56 min Scan# 570
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tot Ion	Ratio	Lower	Upper
96	100		
61	142.2	0.0	284.6
98	64.0	0.0	127.2




#28

Bromochloromethane
 Concen: 47.596 ug/l
 RT: 4.98 min Scan# 639
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion	Ratio	Lower	Upper
49	100		
129	1.6	0.0	4.0
130	78.7	63.0	94.4





#29

Tetrahydrofuran

Concen: 273.186 ug/l

RT: 5.09 min Scan# 656

Delta R.T. -0.01 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 42 Resp: 641442

Ion Ratio Lower Upper

42 100

72 46.0 37.0 55.4




71 42.8 34.1 51.1



Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



#30

Chloroform

Concen: 45.877 ug/l

RT: 5.18 min Scan# 671

Delta R.T. -0.00 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 83 Resp: 349639

Ion Ratio Lower Upper



83 100


85 62.7 52.0 78.0



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0









#31
Cyclohexane
Concen: 46.755 ug/l
RT: 5.55 min Scan# 732
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 56 Resp: 318147
Ion Ratio Lower Upper
56 100
69 30.6 25.7 38.5
84 85.7 70.0 105.0



#32
1,1,1-Trichloroethane
Concen: 45.405 ug/l
RT: 5.46 min Scan# 718
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion: 97 Resp: 313281
Ion Ratio Lower Upper
97 100
99 63.9 51.4 77.2
61 41.8 33.5 50.3









#33
 1,2-Dichloroethane-d4
 Concen: 49.019 ug/l
 RT: 6.03 min Scan# 811
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 65 Resp: 239978
 Ion Ratio Lower Upper
 65 100
 67 53.3 0.0 104.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.83 min Scan# 942
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion: 114 Resp: 644834
 Ion Ratio Lower Upper
 114 100
 63 20.4 0.0 42.6
 88 16.7 0.0 32.8









#35
 Dibromofluoromethane
 Concen: 47.992 ug/l
 RT: 5.47 min Scan# 719
 Delta R.T. 0.01 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion:113 Resp: 190553
 Ion Ratio Lower Upper
 113 100
 111 103.5 83.0 124.6
 192 21.8 17.7 26.5



#36
 1,1-Dichloropropene
 Concen: 43.854 ug/l
 RT: 5.77 min Scan# 768
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion: 75 Resp: 267746
 Ion Ratio Lower Upper
 75 100
 110 35.5 17.4 52.3
 77 30.7 24.2 36.4









#37
Ethyl Acetate
Concen: 48.245 ug/l
RT: 4.79 min Scan# 608
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 43 Resp: 353634
Ion Ratio Lower Upper
43 100
61 14.3 11.3 16.9
70 11.7 9.3 13.9



#38
Carbon Tetrachloride
Concen: 42.482 ug/l
RT: 5.76 min Scan# 766
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion: 117 Resp: 272569
Ion Ratio Lower Upper
117 100
119 98.2 76.3 114.5
121 32.2 24.6 37.0









#39
Methylvlcyclohexane
Concen: 45.379 ug/l
RT: 7.44 min Scan# 1042
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion: 83 Resp: 300929
Ion Ratio Lower Upper
83 100
55 77.4 61.7 92.5
98 49.3 38.7 58.1




Abundance Ion 83.00 (82.70 to 83.70): VX0
Ion 55.00 (54.70 to 55.70): VX0
Ion 98.00 (97.70 to 98.70): VX0






#40
Benzene
Concen: 44.648 ug/l
RT: 6.11 min Scan# 824
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion: 78 Resp: 811771
Ion Ratio Lower Upper
78 100
77 23.5 18.9 28.3



Abundance Ion 78.00 (77.70 to 78.70): VX0
Ion 77.00 (76.70 to 77.70): VX0









#41
Methacrylonitrile
Concen: 46.776 ug/l
RT: 5.00 min Scan# 642
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion:	41	Resp:	188334
Ion	Ratio	Lower	Upper
41	100		
39	54.2	43.0	64.4
67	74.2	58.9	88.3
52	29.3	23.1	34.7



#42
1,2-Dichloroethane
Concen: 43.417 ug/l
RT: 6.17 min Scan# 833
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion:	62	Resp:	280726
Ion	Ratio	Lower	Upper
62	100		
98	10.2	0.0	19.8





#43

Isopropyl Acetate

Concen: 47.580 ug/l

RT: 6.42 min Scan# 874

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 43 Resp: 544001

Ion Ratio Lower Upper

43 100

61 21.6 16.8 25.2




87 13.8 11.2 16.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#44

Trichloroethene

Concen: 41.151 ug/l

RT: 7.19 min Scan# 1001

Delta R.T. -0.00 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 130 Resp: 237385

Ion Ratio Lower Upper



130 100


95 94.3 0.0 198.2



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0





#45

1,2-Dichloropropane

Concen: 45.430 ug/l

RT: 7.49 min Scan# 1050

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 63 Resp: 210275

Ion Ratio Lower Upper




63 100

65 32.3 24.2 36.2



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 64.90 (64.60 to 65.60): VX0



#46

Dibromomethane

Concen: 44.591 ug/l

RT: 7.64 min Scan# 1075

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion: 93 Resp: 138628

Ion Ratio Lower Upper

93 100

95 83.2 66.6 100.0



174 109.2 86.2 129.4




Abundance Ion 92.80 (92.50 to 93.50): VX0

Ion 94.80 (94.50 to 95.50): VX0

Ion 173.70 (173.40 to 174.40): V





#47

Bromodichloromethane

Concen: 44.205 ug/l

RT: 7.88 min Scan# 1114

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 83 Resp: 284558

Ion Ratio Lower Upper

83 100

85 64.3 52.0 78.0




127 8.5 7.4 11.0



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 48.013 ug/l

RT: 7.75 min Scan# 1093

Delta R.T. 0.01 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion: 41 Resp: 262840

Ion Ratio Lower Upper

41 100

69 85.4 68.6 102.8



39 51.3 42.4 63.6




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: 1110.092 ug/l

RT: 7.71 min Scan# 1087

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 88 Resp: 142218

Ion Ratio Lower Upper

88 100

43 29.7 24.2 36.4




58 71.1 56.5 84.7



Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 47.805 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 98 Resp: 738792

Ion Ratio Lower Upper



98 100


100 67.0 53.8 80.6



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0





#51

4-Methyl-2-Pentanone

Concen: 249.097 ug/l

RT: 8.62 min Scan# 1236

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 43 Resp: 1773369

Ion Ratio Lower Upper

43 100

58 41.5 33.0 49.6



#52

Toluene

Concen: 44.105 ug/l

RT: 8.77 min Scan# 1260

Delta R.T. -0.00 min

Lab File: VX016470.D




Acq: 28 May 2020 00:45


Tgt Ion: 92 Resp: 505672

Ion Ratio Lower Upper

92 100

91 168.8 135.0 202.4





#53

t-1,3-Dichloropropene

Concen: 44.735 ug/l

RT: 9.02 min Scan# 1302

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 75 Resp: 327449

Ion Ratio Lower Upper

75 100

77 31.4 25.5 38.3



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 76.90 (76.60 to 77.60): VX0

9.02

250000

200000



150000

100000

50000

0

Time--> 8.95 9.00 9.05 9.10



#54

cis-1,3-Dichloropropene

Concen: 44.703 ug/l

RT: 8.42 min Scan# 1202

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion: 75 Resp: 346350

Ion Ratio Lower Upper

75 100

77 31.6 24.6 36.8

39 45.0 35.3 52.9



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 76.90 (76.60 to 77.60): VX0

Ion 39.00 (38.70 to 39.70): VX0

8.42

300000

250000

200000


150000


100000

50000

0

Time--> 8.30 8.35 8.40 8.45 8.50









#55
 1,1,2-Trichloroethane
 Concen: 44.946 ug/l
 RT: 9.20 min Scan# 1330
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:	97	Resp:	204887
Ion	Ratio	Lower	Upper
97	100		
83	86.0	70.9	106.3
85	56.8	43.4	65.2
99	61.6	51.0	76.6




Abundance Ion 96.90 (96.60 to 97.60): VX0
 Ion 82.90 (82.60 to 83.60): VX0
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 98.90 (98.60 to 99.60): VX0






#56
 Ethyl methacrylate
 Concen: 47.648 ug/l
 RT: 9.16 min Scan# 1324
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion:	69	Resp:	353747
Ion	Ratio	Lower	Upper
69	100		
41	64.6	50.7	76.1
39	31.8	25.8	38.6



Abundance Ion 69.00 (68.70 to 69.70): VX0
 Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0





#57

1,3-Dichloropropane

Concen: 45.280 ug/l

RT: 9.35 min Scan# 1356

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 76 Resp: 354295

Ion Ratio Lower Upper

76 100

78 32.3 26.2 39.4



#58

2-Chloroethyl Vinyl ether

Concen: 235.581 ug/l

RT: 8.29 min Scan# 1182

Delta R.T. -0.00 min

Lab File: VX016470.D




Acq: 28 May 2020 00:45


Tgt Ion: 63 Resp: 912901

Ion Ratio Lower Upper

63 100

106 28.6 23.2 34.8









#59
2-Hexanone
Concen: 255.180 ug/l
RT: 9.48 min Scan# 1376
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 43 Resp: 1412748
Ion Ratio Lower Upper
43 100
58 57.7 28.8 86.4



#60
Dibromochloromethane
Concen: 44.275 ug/l
RT: 9.57 min Scan# 1391
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion: 129 Resp: 227129
Ion Ratio Lower Upper
129 100
127 77.5 38.6 116.0





#61

1,2-Dibromoethane

Concen: 44.973 ug/l

RT: 9.65 min Scan# 1405

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:107 Resp: 219539

Ion Ratio Lower Upper


107 100

109 93.7 75.5 113.3





Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



Time-->



#62

4-Bromofluorobenzene

Concen: 46.265 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion: 95 Resp: 276268

Ion Ratio Lower Upper

95 100

174 80.2 0.0 163.0


176 77.6 0.0 156.8




Abundance Ion 94.90 (94.60 to 95.60): VX0


Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



Time-->





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:117 Resp: 571706

Ion Ratio Lower Upper

117 100

82 57.5 45.0 67.6




119 33.3 25.8 38.8



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#64

Tetrachloroethene

Concen: 39.957 ug/l

RT: 9.32 min Scan# 1350

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Tgt Ion:164 Resp: 255575


Ion Ratio Lower Upper

164 100

166 129.7 100.9 151.3

129 94.0 73.2 109.8

131 91.0 70.8 106.2






Abundance Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V





#65

Chlorobenzene

Concen: 46.000 ug/l

RT: 10.12 min Scan# 1482

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:112 Resp: 536155

Ion Ratio Lower Upper




112 100

114 32.0 26.0 39.0



Abundance Ion 111.90 (111.60 to 112.60): V

Ion 113.90 (113.60 to 114.60): V



#66

1,1,1,2-Tetrachloroethane

Concen: 45.845 ug/l

RT: 10.21 min Scan# 1496

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion:131 Resp: 203972

Ion Ratio Lower Upper

131 100

133 94.4 48.4 145.3



119 63.8 32.0 96.2




Abundance Ion 130.80 (130.50 to 131.50): V

Ion 132.80 (132.50 to 133.50): V

Ion 118.80 (118.50 to 119.50): V









#67
Ethyl Benzene
Concen: 45.991 ug/l
RT: 10.24 min Scan# 1501
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 91 Resp: 954405
Ion Ratio Lower Upper
91 100
106 30.3 25.0 37.4



#68
m/p-Xylenes
Concen: 92.675 ug/l
RT: 10.34 min Scan# 1518
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion: 106 Resp: 713058
Ion Ratio Lower Upper
106 100
91 204.5 163.8 245.8









#69
o-Xylene
Concen: 46.552 ug/l
RT: 10.68 min Scan# 1574
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion:106 Resp: 346881
Ion Ratio Lower Upper
106 100
91 214.9 109.6 328.8



#70
Styrene
Concen: 47.597 ug/l
RT: 10.70 min Scan# 1576
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion:104 Resp: 609625
Ion Ratio Lower Upper
104 100
78 49.2 40.2 60.2
103 53.4 43.1 64.7





#71

Bromoform

Concen: 47.158 ug/l

RT: 10.84 min Scan# 1600

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:173 Resp: 177323

Ion Ratio Lower Upper




173 100

175 49.1 23.9 71.7

254 0.2 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion:152 Resp: 286924

Ion Ratio Lower Upper



152 100


115 77.0 39.9 119.6

150 169.2 0.0 341.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 45.860 ug/l

RT: 11.00 min Scan# 1626

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:105 Resp: 911706

Ion Ratio Lower Upper

105 100

120 26.2 13.1 39.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

800000

600000

400000


200000

0


11.00

10.90

11.10



Time-->



#74

N-ethyl acetate

Concen: 49.912 ug/l

RT: 10.88 min Scan# 1607

Delta R.T. 0.01 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Tgt Ion: 43 Resp: 456418


Ion Ratio Lower Upper

43 100

70 44.8 36.1 54.1

55 24.5 20.2 30.4

61 25.9 20.3 30.5



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

500000

400000

300000

200000


100000

0


10.88

10.80

11.00



Time-->



#75

1,1,2,2-Tetrachloroethane

Concen: 54.157 ug/l

RT: 11.26 min Scan# 1668

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 83 Resp: 269836

Ion Ratio Lower Upper

83 100

131 10.3 5.1 15.3




85 63.8 32.0 96.2



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): VX0

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 62.415 ug/l

RT: 11.28 min Scan# 1672

Delta R.T. -0.00 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 75 Resp: 393633

Ion Ratio Lower Upper



75 100


77 32.4 21.3 63.7



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 45.503 ug/l

RT: 11.24 min Scan# 1665

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:156 Resp: 237266

Ion Ratio Lower Upper

156 100

77 160.6 79.6 238.8




158 96.0 48.1 144.4



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 46.623 ug/l

RT: 11.34 min Scan# 1682

Delta R.T. -0.00 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 91 Resp: 1035529

Ion Ratio Lower Upper



91 100


120 23.1 11.6 34.7



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 45.524 ug/l

RT: 11.40 min Scan# 1692

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 91 Resp: 617190




Ion Ratio Lower Upper

91 100

126 33.6 16.8 50.3



Abundance Ion 91.00 (90.70 to 91.70): VX016449.D (-1688) (-)



#80

1,3,5-Trimethylbenzene

Concen: 45.962 ug/l

RT: 11.49 min Scan# 1707

Delta R.T. 0.01 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 105 Resp: 762968



Ion Ratio Lower Upper


105 100

120 48.7 24.3 72.8



Abundance Ion 105.00 (104.70 to 105.70): VX016449.D (-1699) (-)





#81

trans-1,4-Dichloro-2-butene

Concen: 45.998 ug/l

RT: 11.06 min Scan# 1636

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 75 Resp: 117933

Ion Ratio Lower Upper

75 100

53 88.2 71.3 106.9




89 44.0 36.6 55.0



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#82

4-Chlorotoluene

Concen: 45.782 ug/l

RT: 11.49 min Scan# 1707

Delta R.T. -0.00 min

Lab File: VX016470.D


Acq: 28 May 2020 00:45

Tgt Ion: 91 Resp: 732958

Ion Ratio Lower Upper



91 100


126 29.1 14.7 44.1



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V









#83
 tert-Butylbenzene
 Concen: 48.123 ug/l
 RT: 11.76 min Scan# 1750
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion	Ratio	Lower	Upper
119	100		
91	60.7	31.3	93.8
134	23.6	12.0	36.0



#84
 1,2,4-Trimethylbenzene
 Concen: 46.112 ug/l
 RT: 11.79 min Scan# 1756
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion	Ratio	Lower	Upper
105	100		
120	45.4	22.6	67.7









#85
 sec-Butylbenzene
 Concen: 47.247 ug/l
 RT: 11.93 min Scan# 1779
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion:105 Resp: 840670
 Ion Ratio Lower Upper
 105 100
 134 19.8 10.1 30.3



#86
 p-Isopropyltoluene
 Concen: 46.555 ug/l
 RT: 12.05 min Scan# 1798
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion:119 Resp: 775949
 Ion Ratio Lower Upper
 119 100
 134 25.8 12.7 38.0
 91 23.3 11.7 35.0





#87

1,3-Dichlorobenzene

Concen: 45.289 ug/l

RT: 12.01 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:146 Resp: 409121

Ion Ratio Lower Upper

146 100

111 40.8 20.0 60.0




148 64.2 31.9 95.5



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#88

1,4-Dichlorobenzene

Concen: 44.110 ug/l

RT: 12.08 min Scan# 1803

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion:146 Resp: 407533

Ion Ratio Lower Upper

146 100

111 40.3 19.4 58.2



148 64.2 31.4 94.3




Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V









#89
n-Butylbenzene
Concen: 47.592 ug/l
RT: 12.37 min Scan# 1851
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 91 Resp: 660710
Ion Ratio Lower Upper
91 100
92 55.1 27.2 81.5
134 26.2 13.1 39.3



#90
Hexachloroethane
Concen: 47.212 ug/l
RT: 12.58 min Scan# 1885
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion: 117 Resp: 135920
Ion Ratio Lower Upper
117 100
201 83.1 43.1 129.3









#91
 1,2-Dichlorobenzene
 Concen: 45.063 ug/l
 RT: 12.38 min Scan# 1852
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion	Ratio	Lower	Upper
146	100		
111	42.2	21.4	64.3
148	64.0	31.9	95.7



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 50.992 ug/l
 RT: 12.98 min Scan# 1951
 Delta R.T. -0.00 min
 Lab File: VX016470.D
 Acq: 28 May 2020 00:45

Tgt Ion	Ratio	Lower	Upper
75	100		
155	87.5	44.1	132.4
157	113.6	57.8	173.3





#93

1,2,4-Trichlorobenzene

Concen: 47.236 ug/l

RT: 13.63 min Scan# 2057

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:180 Resp: 274563

Ion Ratio Lower Upper

180 100

182 97.4 48.3 144.8




145 30.8 15.4 46.2



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V



#94

Hexachlorobutadiene

Concen: 46.459 ug/l

RT: 13.77 min Scan# 2080

Delta R.T. -0.00 min

Lab File: VX016470.D

Acq: 28 May 2020 00:45


Tgt Ion:225 Resp: 106027

Ion Ratio Lower Upper

225 100

223 62.2 32.3 96.8



227 63.5 31.6 94.7




Abundance Ion 224.70 (224.40 to 225.40): V

Ion 222.70 (222.40 to 223.40): V

Ion 226.70 (226.40 to 227.40): V









#95
Naphthalene
Concen: 49.383 ug/l
RT: 13.82 min Scan# 2088
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion:128 Resp: 984027
Ion Ratio Lower Upper
128 100
127 12.9 10.6 15.8
129 11.0 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 47.983 ug/l
RT: 14.01 min Scan# 2119
Delta R.T. -0.00 min
Lab File: VX016470.D
Acq: 28 May 2020 00:45

Tgt Ion:180 Resp: 273242
Ion Ratio Lower Upper
180 100
182 96.7 48.6 145.8
145 31.3 16.6 49.8

