

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052720\
 Data File : VX016461.D
 Acq On : 27 May 2020 21:24
 Operator : JC/SP
 Sample : PB129126TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129126TB

Quant Time: May 28 07:37:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	276144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	451509	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	438119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	219693	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	166741	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
35) Dibromofluoromethane	5.47	113	131810	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
50) Toluene-d8	8.70	98	561063	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
62) 4-Bromofluorobenzene	11.12	95	223346	53.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.84%	
Target Compounds						
16) Acetone	2.42	43	22606	15.046	ug/l	98
18) Methyl Acetate	2.75	43	4122	1.070	ug/l	95
20) Methylene Chloride	2.83	84	5768	1.856	ug/l	87
43) Isopropyl Acetate	6.42	43	89409	11.168	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052720\
Data File : VX016461.D
Acq On : 27 May 2020 21:24
Operator : JC/SP
Sample : PB129126TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129126TB

Quant Time: May 28 07:37:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

