

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052720\
 Data File : VX016462.D
 Acq On : 27 May 2020 21:46
 Operator : JC/SP
 Sample : L2767-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1005

Quant Time: May 28 05:34:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	275396	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	446677	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	454988	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	234302	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	167989	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
35) Dibromofluoromethane	5.47	113	133257	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
50) Toluene-d8	8.70	98	562641	52.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.12%	
62) 4-Bromofluorobenzene	11.12	95	229350	55.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.90%	
Target Compounds						
8) Diethyl Ether	2.17	74	3875	2.043	ug/l	96
16) Acetone	2.42	43	31603	21.091	ug/l	98
18) Methyl Acetate	2.75	43	4127	1.074	ug/l	95
20) Methylene Chloride	2.84	84	6635	2.140	ug/l #	86
43) Isopropyl Acetate	6.41	43	78110	9.862	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052720\
Data File : VX016462.D
Acq On : 27 May 2020 21:46
Operator : JC/SP
Sample : L2767-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1005

Quant Time: May 28 05:34:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

