

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052720\
 Data File : VX016464.D
 Acq On : 27 May 2020 22:32
 Operator : JC/SP
 Sample : L2757-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0AP3

Quant Time: May 28 07:49:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	284214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	465088	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	460413	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	252178	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.04	65	172623	48.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.82%	
35) Dibromofluoromethane	5.47	113	137799	48.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.24%	
50) Toluene-d8	8.70	98	579342	51.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.96%	
62) 4-Bromofluorobenzene	11.12	95	259815	60.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.64%	

Target Compounds

						Qvalue
16) Acetone	2.42	43	183057	118.377	ug/l	96
18) Methyl Acetate	2.75	43	343289	86.554	ug/l	100
20) Methylene Chloride	2.84	84	6378	1.994	ug/l	98
25) 2-Butanone	4.64	43	143470	54.753	ug/l	99
29) Tetrahydrofuran	5.10	42	35435	20.932	ug/l	99
40) Benzene	6.12	78	143710	10.959	ug/l	99
43) Isopropyl Acetate	6.42	43	90082	10.924	ug/l	97
52) Toluene	8.77	92	767720	92.841	ug/l	100
59) 2-Hexanone	9.48	43	247162	61.898	ug/l	95
67) Ethyl Benzene	10.24	91	571915	34.222	ug/l	100
68) m/p-Xylenes	10.34	106	558524	90.138	ug/l	99
69) o-Xylene	10.68	106	399684	66.604	ug/l	99
70) Styrene	10.70	104	97061	9.410	ug/l #	40
73) Isopropylbenzene	11.00	105	36631	2.096	ug/l	99
78) n-propylbenzene	11.34	91	114154	5.848	ug/l	99
80) 1,3,5-Trimethylbenzene	11.49	105	131632	9.022	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	522890	35.439	ug/l	99
89) n-Butylbenzene	12.37	91	30626	2.510	ug/l #	59
95) Naphthalene	13.82	128	1569812	89.635	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052720\
Data File : VX016464.D
Acq On : 27 May 2020 22:32
Operator : JC/SP
Sample : L2757-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AP3

Quant Time: May 28 07:49:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

