

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052720\
 Data File : VX016468.D
 Acq On : 28 May 2020 00:02
 Operator : JC/SP
 Sample : L2786-02 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LG-C

Quant Time: May 28 05:05:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	352330	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	568643	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	578513	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	298208	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	201947	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
35) Dibromofluoromethane	5.47	113	162723	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
50) Toluene-d8	8.70	98	708261	51.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	11.12	95	296009	56.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.42%	
Target Compounds						
16) Acetone	2.42	43	19346	10.092	ug/l	99
17) Carbon Disulfide	2.55	76	1168665	116.082	ug/l	99
68) m/p-Xylenes	10.34	106	2070	0.266	ug/l	99
95) Naphthalene	13.82	128	23134	1.117	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052720\
Data File : VX016468.D
Acq On : 28 May 2020 00:02
Operator : JC/SP
Sample : L2786-02 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LG-C

Quant Time: May 28 05:05:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

