

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022191.D
 Acq On : 26 May 2021 12:46
 Operator : JC/MD
 Sample : M2480-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CYCLE-WATER

Quant Time: May 26 13:12:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	68582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	142177	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124067	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44956	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56162	53.09	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.18%
35) Dibromofluoromethane	5.404	113	43915	48.71	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.42%
50) Toluene-d8	8.653	98	173394	49.92	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.84%
62) 4-Bromofluorobenzene	11.085	95	58679	45.71	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.42%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	37724	81.89	ug/l #	88
30) Chloroform	5.111	83	9822	5.83	ug/l	93
47) Bromodichloromethane	7.830	83	1886	1.35	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
Data File : VX022191.D
Acq On : 26 May 2021 12:46
Operator : JC/MD
Sample : M2480-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CYCLE-WATER

Quant Time: May 26 13:12:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

