

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022196.D
 Acq On : 26 May 2021 14:43
 Operator : JC/MD
 Sample : M2526-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20210525

Quant Time: May 27 04:05:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	131811	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	115701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43153	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53347	53.85	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.70%	
35) Dibromofluoromethane	5.397	113	41200	49.29	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.58%	
50) Toluene-d8	8.653	98	160256	49.77	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.54%	
62) 4-Bromofluorobenzene	11.085	95	54785	46.03	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.06%	
Target Compounds						
95) Naphthalene	13.780	128	7429	2.20	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
Data File : VX022196.D
Acq On : 26 May 2021 14:43
Operator : JC/MD
Sample : M2526-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20210525

Quant Time: May 27 04:05:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

