

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022197.D
 Acq On : 26 May 2021 15:07
 Operator : JC/MD
 Sample : M2526-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20210525

Quant Time: May 27 04:06:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	66020	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	136452	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	119601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46801	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	53077	52.12	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.24%
35) Dibromofluoromethane	5.404	113	41758	48.26	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.52%
50) Toluene-d8	8.653	98	165453	49.64	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.28%
62) 4-Bromofluorobenzene	11.085	95	59136	48.00	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.00%
Target Compounds						
16) Acetone	2.386	43	4169	9.40	ug/l	Qvalue 100
20) Methylene Chloride	2.794	84	2757	2.80	ug/l	# 90
95) Naphthalene	13.780	128	3990	1.09	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
Data File : VX022197.D
Acq On : 26 May 2021 15:07
Operator : JC/MD
Sample : M2526-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20210525

Quant Time: May 27 04:06:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

