

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
 Data File : VX022198.D
 Acq On : 26 May 2021 15:31
 Operator : JC/MD
 Sample : M2520-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-03-2021-05-24DL

Manual Integrations
 APPROVED

MMDadoda
 5/28/2021 2:34:27 PM

Quant Time: May 27 04:06:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	130341	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	115949	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49388	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50299	52.70	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.40%			
35) Dibromofluoromethane	5.397	113	40486	48.98	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.96%			
50) Toluene-d8	8.653	98	160796	50.50	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.00%			
62) 4-Bromofluorobenzene	11.085	95	57303	48.69	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.38%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	3679	8.85	ug/l	# 86
40) Benzene	6.050	78	32012	8.63	ug/l	95
52) Toluene	8.720	92	40848	17.39	ug/l	97
67) Ethyl Benzene	10.201	91	202885	45.69	ug/l	99
68) m/p-Xylenes	10.305	106	40237	23.98	ug/l	97
69) o-Xylene	10.646	106	29887	18.31	ug/l	98
73) Isopropylbenzene	10.963	105	14097	3.37	ug/l	96
78) n-propylbenzene	11.305	91	6656	1.36	ug/l	100
80) 1,3,5-Trimethylbenzene	11.457	105	14000	4.04	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	54507	15.63	ug/l	99
86) p-Isopropyltoluene	12.012	119	3230m	0.91	ug/l	
95) Naphthalene	13.780	128	2234951	578.61	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052721\
Data File : VX022198.D
Acq On : 26 May 2021 15:31
Operator : JC/MD
Sample : M2520-03DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-03-2021-05-24DL

Manual Integrations
APPROVED

MMDadoda
5/28/2021 2:34:27 PM

Quant Time: May 27 04:06:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

